pipetmaX® Maximize Reproducibility of Your Biological Sample Prep



SPEC SHEET | LIQUID HANDLING

PIPETMAX[®]

Maximize pace and capabilities with the ultimate lab assistant

Maximize consistency in routine pipetting tasks

Maximize reproducibility in biological assays



PIPETMAX Versatility

- Ideal for PCR, qPCR, cell-based assays, NGS prep, ELISA prep, and tip-based sample prep all on the same PIPETMAX
- Flexible settings for various sample types, including aqueous, cells, tissue cultures, and biological fluids
- Large working volume dynamic range of 1–1200 μL
- Many types of labware and devices: 96- and 384-well microplates, strip plates, and cell culture formats
- PIPETMAN® standards in reproducibility, precision, and accuracy
- Easy configuration and run with a simple, user-friendly interface



Free yourself from tedious sample prep work Let PIPETMAX prepare samples

for you! Enjoy more freedom away from the bench and focus on what's important—your next discovery.

A trul Config reager The ha to be a of you

A truly configurable system Configure your runs using any reagent and any protocol you want. The hardware and software are built to be adaptable to the unique needs of your research.





your pipetting to be reliable and consistent—plate to plate, lot to lot, time after time.



Versatile

Your choice of hardware to fit your needs and pipette heads can be calibrated like PIPETMAN.



TECHNICAL SPECIFICATIONS - PIPETMAX® 268

							Pipette Head	PIPETMAN® DIAMOND Tips	Pipette Head Volume Range (µL)	Flow Rate Range (mL/min)
MAXIMUM PERMISSIBLE ERRORS						O Tips		DL10	1-20	
Pipette Head	Volume (µL)	Systematic Error (µL)	Random Error (µL)	Systematic Error (%)	Random Error (%)	DIAMOND some tips)	MAX8x20	DFL10	1-10	- 0.0225-3.75
MAX8x20	1	±0.08	≤ 0.05	8.00	5.00	DIA			1-10	
	10	±0.15	≤ 0.10	1.50	1.00	1AN [®] le for s		DF30	1-20	
	20	±0.25	≤ 0.12	1.25	0.60	ETN ailabl		D 200	1.00	
MAX8x200	20	±0.50	≤ 0.16	2.50	0.80	a PIF		D200	1-20	
	100	±1.00	≤ 0.30	1.00	0.30	d optic	MAX8x200	D200	20-200	
	200	±2.00	≤ 0.50	1.00	0.25	lead				0.225-37.5
MAX4x1200	50	±4.00	≤ 0.70	8.00	1.40	Pipette Heads and PIPETMAN [*] (Sterilized option available for		DF200	20-200	
	120	±4.00	≤ 0.70	3.33	0.58		MAX4x1200	D1200	50-1200	
	600	±6.00	≤ 1.50	1.00	0.25					1.125-187.5
	1200	±9.60	≤ 1.80	0.80	0.15			DF1200	50-1200	

Communication	USB				
Connections	Three USB host ports and one USB device port Two inputs (contact closure, TTL), two relay outputs, and one switched ±12V DC 1A output NOTICE: Switching voltages higher than 30V or greater than 1A of current may damage the instrument				
Control	Touchscreen tablet, or PC control via USB and TRILUTION® micro software				
Dimensions (W x D x H)	PIPETMAX with rotating cover 54.4 x 65.5 x 53.1 cm (21.4 x 25.8 x 20.9 in.) PIPETMAX with rotating cover installed on optional riser assembly for off-bed tip disposal 54.4 x 65.5 x 69.6 cm (21.4 x 25.8 x 27.4 in.) PIPETMAX without rotating cover 50.8 x 64.3 x 49.5 cm (20 x 25.3 x 19.5 in.) PIPETMAX without rotating cover installed on optional riser assembly for off-bed tip disposal 52.3 x 65 x 65.8 cm (20.6 x 25.6 x 25.9 in.)				
Environmental conditions	Altitude: up to 2000 m Temperature range: 5°C-40°C Humidity: Maximum relative humidity 80% for temperatures up to 31°C, decreasing linearly to 50% relative humidity at 40°C				
Front Panel	Two USB host ports and STOP button				
	DESCRIPTION	MATERIAL			
	Tips	100% Virgin Polypropylene			
	Tip Disposal Bin	Pro-fax 6523 (Polypropylene with colorant)			
Liquid Contact Materials	Tip Disposal Bin (Off-bed)	Polypropylene			
	Tip Chute	Pro-fax 6523 (Polypropylene with colorant), Aluminum 50 with Hentzen, URA-ZEN, White, Matte Finish, Fine Texture Paint			
Power Requirements - External Power Supply	Voltage Input: Frequency: 50 to 60Hz; Voltage: 100-240V AC Voltage Output: Voltage: 24V DC; Current Rating: 6.25A, 150W				
Removable Tray Capacity	9-position removable tray (microplate footprints, but not for 384-well microplates) 9-position removable tray for 384-well microplates				
Safety and Compliance	PIPETMAX has been certified to safety standards specified for Canada, Europe, and the United States. Refer to the instrument rear panel label and the Declaration of Conformity document for the current standards to which the instrument has been found compliant.				