



CERTIFICATE OF ACCREDITATION

ANSI-ASQ National Accreditation Board

500 Montgomery Street, Suite 625, Alexandria, VA 22314, 877-344-3044

This is to certify that

Gilson, Inc.
3101 Laura Lane, Suite 100
Middleton, WI 53562

has been assessed by ANAB
and meets the requirements of international standard

ISO/IEC 17025:2005

while demonstrating technical competence in the fields of

CALIBRATION

Refer to the accompanying Scope of Accreditation for information regarding the types of calibrations and/or tests to which this accreditation applies.

AC-1731

Certificate Number



ANAB Approval

Certificate Valid: 07/18/2018-09/28/2020
Version No. 006 Issued: 07/18/2018



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

Gilson, Inc.

3101 Laura Lane, Suite 100
Middleton, WI 53562
Michael Zwart 800-445-7661

CALIBRATION

Valid to: September 28, 2020

Certificate Number: AC-1731

Mass and Mass Related

Table with 4 columns: Parameter / Equipment, Range, Expanded Uncertainty of Measurement (+/-), Reference Standard, Method and/or Equipment. Row 1: Pipette, (0.2 to 2) µL to (5 000 to 10 000) µL, 0.89% to 0.11% of reading, ISO 8655-6; Gravimetric method.

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 (k=2), corresponding to a confidence level of approximately 95%.

Notes:

- 1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-1731.

Signature of R. D. Zwart
Vice President

